



TECHNICAL DATA

52120A Transconductance Amplifier





Key features

- Supply DC current up to 100 amps and AC current up to 120 amps, with accuracies to 140 ppm.
- Generate 3,000 or 6,000 amps using accessory coils, up to 360 amps with three 52120As are connected in parallel.
- Frequency: DC to 10 kHz, a burden voltage (compliance): 4.5 V @ 120 A, inductive drive capability: 1 mH load.
- Allows for parallel operation with 2 or 3 amplifiers up to 360 A in a single phase, or up to 10 amplifiers and 1200 A in standalone mode.
- GPIB remote operation

Product overview: 52120A Transconductance Amplifier

The 52120A is designed for users whose ability to address their calibration workload may be limited by the output current, accuracy and drive capability of their present test equipment, including:

- Calibration professionals in a calibration/standards lab or an electrical utility
- Manufacturers of power/energy instrumentation and meters, power quality analyzers or power converters
- Users of electrical test and measurement equipment

The 52120A operates as a transconductance amplifier with:

- 5522A/5502A/5520A/5500A Multi-Product Calibrators
- 5730A*/5700A/5720A Multifunction Calibrator
- 6105A*/6100B* Electrical Power Standard
- 5080A Multi-Product Calibrator
- 9100 Universal Calibration System
- Any calibrator, signal generator or power supply capable of sourcing 2 V or 200 mA, dc or ac

*Special closed-loop operating mode

You may also operate your 52120A in closed-loop mode, seamlessly communicating with your Fluke Calibration 6105A or 6100B Electrical Power Standard to deliver enhanced 52120A accuracy.

Specifications: 52120A Transconductance Amplifier

Operating limits	
Output ranges	2 A, 20 A, 120 A (100 A dc)
Input current range	200 mA, 200 mA, 120 mA (100 mA dc)
Current gain	10, 100, 1000
Input voltage range	2 V, 2 V, 1.2 V (1.0 V dc)
Transconductance	1,10,100 Siemens
Frequency	To 10 kHz
Maximum output compliance voltage	4.5 Vrms (6.4 Vpeak)
Inductive drive capability:	1 mH load

Output isolation, current terminal to earth	600 Vrms, 850 Vpeak, dc to 850 Hz
Performance specifications	
AC accuracy, closed loop with 6105A	To 140 ppm
DC accuracy, standalone	To 160 ppm
AC accuracy, standalone	To 350 ppm
Phase angle accuracy	0.006°
Load dependent phase angle shift	<0.001° @ 60 Hz
General specifications	
Input line voltage range	100 V to 240 V with up to ±10 % fluctuations
Transient overvoltage	Impulse withstand (overvoltage); Category II of IEC 60364-4-443
Frequency	47 Hz to 63 Hz
Maximum consumption	< 1500 VA
Calibration Documentation	Report of calibration with data included; 17025 accredited report optional
Dimensions with feet (H x W x L)	192 mm x 432 mm x 645 mm (7.6 inches x 17 inches x 25.5 inches)
Dimensions without feet (H x W x L)	178 mm x 432 mm x 645 mm (7 inches x 17 inches x 25.5 inches)
Weight	25 kg (54 lb.)
Design standards and compliance	Designed to EN 61010-1: 2010, CAN/CSA 22.2 No 61010.1-04, ANSI/UL 61010-1:2004, EN 61326-1:2006 CE marked, CSA listed
Operating temperature	5 °C to 35 °C
Calibration temperature range	16 °C to 30 °C
Warm up time	1 hour
Safe operating max. relative humidity (non-condensing)	<80 % 5 °C to 31 °C ramping linearly down to 50 % at 35 °C
Operating altitude	0 m to 2,500 m
Shock and vibration	MIL-PRF-28800F class 3
Specification confidence level	99%

Ordering information



52120A

Transconductance Amplifier

Fluke. *Keeping your world up and running.®*

Fluke Corporation

PO Box 9090, Everett, WA 98206 U.S.A.

For more information call:

In the U.S.A. (800) 443-5853

In Europe/M-East/Africa

+31 (0)40 267 5100

In Canada (800)-36-FLUKE

From other countries +1 (425) 446-5500

BUT. FLUKE SOUTH EAST ASIA PTE LTD

Menera Satu Sentra Kelapa Gading #06-05

Jl. Bulevar Kelapa Gading Kav. LA# No. 1

Summarecon Kelapa Gading

Jakarta Utara 14240

Indonesia

Tel: +62 21 2938 5922

Email: info.asean@fluke.com

www.fluke.com/id-en

©2025 Fluke Corporation. Specifications subject to change without notice.

04/2025

Modification of this document is not permitted without written permission from Fluke Corporation.