

TECHNICAL DATA

Fluke 301E Series AC/DC Clamp Meter





Key features

- The body weighs only 154g and is 16mm thick, which is light and portable and can be put into a pocket
- The slim and thin jaw of only 10mm can easily fit into tightly packed wires
- 1000A AC/DC (E model), 0.01A high resolution
- Tests voltage, resistance and continuity
- Tests current frequency and voltage frequency
- Tests capacitance and diodes
- Backlight

Product overview: Fluke 301E Series AC/DC Clamp Meter

The Fluke 301E Series AC/DC Clamp Meter is designed with a lightweight and compact body, which allows you to carry it around in your pocket. The slim and thin jaw can easily fit into tightly packed wires. The Fluke 301E test current, voltage, resistance, continuity, frequency (voltage and current), capacitance, and diode, etc.. You can handle multiple test needs with just one tool. With the AC/DC function, it can be applied to more application scenarios requiring DC testing, such as new energy of PV and wind power, new energy vehicles, communications, data centers, etc..

Specifications: Fluke 301E Series AC/DC Clamp Meter

Models		301D		301E		
AC Current (45 to 400 Hz)	Range	60.00 A	600.0 A	60.00 A	600.0 A	1000 A
	Resolution	0.01 A	0.1 A	0.01 A	0.1 A	1 A
	Accuracy	2% + 10	2% + 5	2% + 10	2% + 5	2% + 5
DC Current	Range	60.00 A	600.0 A	60.00 A	600.0 A	1000 A
	Resolution	0.01 A	0.1 A	0.01 A	0.1 A	1 A
	Accuracy	2% + 10	2% + 5	2% + 10	2% + 5	2% + 5
AC Voltage (45 to 400 Hz)	Range	600.0 V		600.0 V		
	Resolution	0.1 V		0.1 V		
	Accuracy	1.5% + 5		1.5% + 5		
DC Voltage	Range	600.0 V		600.0 V		
	Resolution	0.1 V		0.1 V		
	Accuracy	1% + 5		1% + 5		
Resistance	Range	600.0 Ω 6.000 kΩ 60.00 kΩ		600.0 Ω 6.000 kΩ 60.00 kΩ		
	Accuracy	1% + 5		1% + 5		

Voltage Frequency	Range	9.999 Hz 99.99 Hz 999.9 Hz 9.999 kHz 99.99 kHz		9.999 Hz 99.99 Hz 999.9 Hz 9.999 kHz 99.99 kHz	
	Accuracy	0.1% + 3		0.1% + 3	
Current Frequency	Range	45.0-400.0 Hz		45.0-400.0 Hz	
	Accuracy	0.1% + 3		0.1% + 3	
Capacitance	Range	9.999 uF	99.99 uF 999.9 uF	9.999 uF	99.99 uF 999.9 uF
	Accuracy	2% + 5	5% + 5	2% + 5	5% + 5
Diode	Range	3.000 V		3.000 V	
	Accuracy	10%		10%	
T-rms Value		-		-	
Continuity		□		□	
Hold		□		□	
Backlight		□		□	
Safety rating		CAT III 300 V		CAT III 300 V	
Weight		154 g		154 g	
Size		190 mm x 52 mm x 16 mm		190 mm x 52 mm x 16 mm	
Jaw Opening		34 mm		34 mm	
Clamp Arm Size		10 mm X 10 mm		10 mm X 10 mm	
Baery		(2) AAA baeries		(2) AAA baeries	
Warranty		1 year		1 year	

General Specifications

Maximum Voltage between any Terminal and Earth Ground	300 V
Maximum Differential Voltage between V/Ω and COM Terminals	600 V
Display (LCD)	6000 counts, updates 3 times per second
Baery Type	2 AAA, IEC LR03
Baery Life	500 hours minimum
Auto Power Off Time	20 minutes
Temperature	Operating: -10 °C to 50 °C □
	Storage: -30 °C to 60 °C
Relative Humidity	Operating Humidity: Non condensing (<10°C); □ 90 % RH at 10 °C to 30 °C; □ 75 % RH at 30 °C to 40 °C; □ 45 % RH at 40 °C to 50 °C;
Altitude	Operating: 2000 m; Storage: 12000 m

Temperature Coefficient	0.1 x (specified accuracy) /°C (<18 °C or >28 °C)
Dimensions (H x W x L)	190 mm x 52 mm x 16 mm
Weight	132 g (301D/E 154 g)
Jaw Opening	34 mm
Ingress Protection Rating	IP30
Safety	IEC 61010--1, IEC61010-2-032 CAT III 300 V, Pollution Degree 2
Electromagnetic Environment	IEC 61326--1: Portable

Ordering information



Fluke 301E Series AC/DC Clamp Meter - EN

Model: Fluke 301E Series AC/DC Clamp Meter

Fluke 301E

- Fluke 301E Clamp Meter
 - TL75 test lead
 - (2) AAA batteries
-

Fluke. *Keeping your world up and running.®*

Fluke Corporation

PO Box 9090, Everett, WA 98206 U.S.A.

For more information call:

In the U.S.A. (800) 443-5853

In Europe/M-East/Africa

+31 (0)40 267 5100

In Canada (800)-36-FLUKE

From other countries +1 (425) 446-5500

BUT. FLUKE SOUTH EAST ASIA PTE LTD

Menera Satu Sentra Kelapa Gading #06-05

Jl. Bulevar Kelapa Gading Kav. LA# No. 1

Summarecon Kelapa Gading

Jakarta Utara 14240

Indonesia

Tel: +62 21 2938 5922

Email: info.asean@fluke.com

www.fluke.com/id-en

©2025 Fluke Corporation. Specifications subject to change without notice.

04/2025

Modification of this document is not permitted without written permission from Fluke Corporation.