

**TECHNICAL DATA** 

# 9940 IO Toolkit v1.0

	×		
Command			
Send	1524 Probe Sett	ings	×
Response	Channel 1 Channe	12	-1
FLUKE,1524,9762234,1.00			
	Probe Type	Serial	
	PBT	▼ 1234	Open
	1		
	Conversion Type	e Calibration Date	,     -
	ITS-90	7 /12/2011	
Ready		_ ,	Save
	Parameters		
	Name	Value	<u> </u>
	RTPW	100.0124566	
	A	-2.545465465e-04	_
	В	5.324325626e-05	
	C	0.0000000000000e+00	
	D	0.0000000000000e+00	
	A4 B4	-5.262762545e-04	Upload
	84	4.635435354e-06	
😪 IOToolkit - 1524 s/n 9762234			X
TIDN?			Cancel
	eady		

## **Key features**

- Interfaces with select Fluke Calibration temperature calibration products via RS-232.
- Unique feature set per instrument.
- Supports new 1523 and 1524 Reference Thermometers.
- Free software, no serial number required.

## Product overview: 9940 IO Toolkit v1.0

## **Communication basics**

Connect the instrument to an available RS-232 port and start the IO Toolkit software. IO Toolkit uses an auto-detection process to establish communications with the connected instrument. Connect the instrument to an available RS-232 port and start the software. The auto-detection process sends identification commands over a range of possible baud rates until a response is received. By default the software attempts communications on COM1. If the serial port being used is not COM1 the auto-detect may be aborted. Otherwise, the auto-detect process will finish and indicate that an instrument could not be detected. The correct port may then be specified.



# Log data

This feature allows data to be downloaded that has been logged by the instrument. Instruments may support two types of logging: auto and demand. IO Toolkit supports both options. Just click the Log Data icon to download logged data.

Auto log data is generally associated with a tag. You can download data for all tags or for a specific tag. Store downloaded data in a single file or separate the data into files by tag.

Demand log data is generally associated with an index. You can download data for all indexes or for a specific index. Downloaded data is stored in a single file.

The log data dialog allows you to browse and select a folder to save the logged data files to. Unless otherwise specified, the default folder is used.

The Download button retrieves and saves data from the indicated instrument log based on the selections made on the dialog.

The Erase button clears data for the indicated log. Auto log data can be erased for all tags or just a single tag. When erasing the demand log all data is erased.

## Log tags

This feature provides a way to manage log tags. Log tags help you manage log data stored in the instrument. To manage log tags, click on the Log Tags icon.

The Log Tags dialog displays the tags that are currently stored in the instrument. Changes made on the dialog are not written to the instrument until the Upload button has been clicked.

You can modify a tag by selecting the tag from the list and modifying the text that appears in the Update tag box. Clicking Apply updates the Tags list with the changes.

To apply the changes in the Tags list to the instrument, click Upload.

## **Probe settings**

This feature enables you to access the probe settings. Probe settings may include information such as probe type, conversion, and calibration information.

The Probe Settings dialog displays the information currently stored for the probe(s). If the instrument has more than one channel settings for the probes are displayed on the corresponding tabs. Changes made on the dialog are not written to the instrument until the Upload button has been clicked.

Settings may be modified using the dialog. Once the changes have been made, click the Upload button to write the changes to the instrument. If the probe settings are protected, a dialog will appear, prompting for a password. Changes will not be written to the instrument unless a valid password is supplied.

Probe configurations can be saved and loaded from the dialog. Saving a probe configuration creates a file containing



the current setup for the channel selected on the dialog. Loading a probe configuration populates the selected channel on the dialog with the stored settings from the selected file.

#### Commands

This feature allows for direct serial communication with a connected instrument. Communication settings are set up automatically during the auto-detection process.

The Commands dialog is similar to other terminal programs. Type supported serial communication commands in the Command box. Click the Send button to send the command to the instrument. Responses received from the instrument are displayed in the Response box.

The information displayed in the Response box can be saved to a text file by clicking the Save button.

This feature is available regardless of the instrument and appears whether or not an instrument is detected.

#### **Settings**

The Settings dialog contains settings and configuration information used by the software.

The COM Port used by the software may be specified on the Communication tab. This option stores the port that the software attempts to detect an instrument on.

The Default Folder Path option designates a folder to use as the default folder for saving and opening files on other dialogs.

Click OK to save changes made on the Settings dialog.

This feature is available regardless of the instrument and appears whether or not an instrument is detected.

The 9940 IO Toolkit ships with <u>supported instruments</u> and provides a way to interface with certain Fluke Calibration temperature calibration products. It uses supported instruments' RS-232 interface to provide access to various functions of the instrument.

Each instrument has a unique feature set, so the same features may not be available for all instruments. The software provides access to those features that are supported by the connected instrument.

This version includes support for the new 1523 and 1524 Reference Thermometers.

This software is free and does not require a serial number to be installed.

Size: 7.0MB

#### Questions and answers about 9940 IO Toolkit software



#### Will the IO Toolkit software allow me to connect to more than one instrument simultaneously from my PC?

A. The IO Toolkit was designed to connect to only one instrument at a time. Running more than one instance of the software simultaneously on a computer has not been tested.

#### The IO Toolkit software does not detect my instrument. What's wrong?

A. The IO Toolkit attempts to establish communications on a single COM port. Verify the COM port being used is the one the instrument is connected to. By default the software is configured to scan COM1. Use the Settings dialog to specify the desired COM port and then try again. Also, make sure the cable being used is a "Null Modem" cable (TX and RX lines cross). You can use HyperTerminal (Windows® 2000/XP/Vista) to verify that the instrument is communicating with the computer.

# The auto-detect process goes through too many baud rates before detecting my instrument. Can I speed this process up?

A. Yes. On the Settings dialog you can specify a default baud rate. The software will start with this baud rate when attempting to auto-detect an instrument. If no instrument is detected the process will continue to try other baud rates.



## **Ordering information**

Command	Send	× 1524 Probe Sett	ings		×
Response		Channel 1 Channe	12		
FLUKE,1524	,9762234,1.00				
		Probe Type		Serial	Open
		PRT	-	1234	
		Conversion Type		Calibration Date	
		ITS-90		7 /12/2011	
Ready		113-30		1///2/2011	Save
roomyss		Parameters			
(© 101 soli	it - 1524 s∕n 9762234	Name RTPw/ A B C D A4 B4	Value 100.0124566 -2.545465465e 5.32432562ee 0.0000000000 0.0000000000 -5.262762545e 4.635435354ee	05 0000e+00 0000e+00 -04	Upload
					Cancel

#### 9940 IO Toolkit v1.0

Connect the instrument to an available RS-232 port and start the IO Toolkit soft Ware.



#### Fluke. Keeping your world up and running.®

Fluke (UK) Ltd. 52 Hurricane Way Norwich, Norfolk NR6 6JB United Kingdom Tel.: +44 (0)20 7942 0708 E-mail: cs.uk@fluke.com www.fluke.com ©2025 Fluke Corporation. All rights reserved. Data subject to alteration without notice. 04/2025

Modification of this document is not permitted without written permission from Fluke Corporation.