

TECHNICAL DATA

# Fluke 802EN Vibration Meter





## Key features

- Three test modes to meet complex test environment in industrial field
- High repeatability to ensure accuracy of continuous testing
- One-button measurement to get assessment result in 6 seconds
- 99 sets of data storage, review in real-time

## Product overview: Fluke 802EN Vibration Meter

Fluke 802EN Vibration Meter

One-button vibration assessment of mechanical equipment

## Specifications: Fluke 802EN Vibration Meter

Inteal Piezoelectric Accelerometer		
Sensitivity (typical)	10.00 mV / m/s <sup>2</sup> ± 5% (1 g = 9.8 m/s <sup>2</sup> )	
Measurement		
Acceleration	Range	0.1 m/s <sup>2</sup> to 199.9 m/s <sup>2</sup> peak (rms √2)
	Resolution	0.1 m/s <sup>2</sup>
	Accuracy	± (5 % + 2 dgt) (LO: 10 Hz to 1 kHz) ± (10 % + 5 dgt) (HI: 1 kHz to 10 kHz)
Velocity	Range	0.1 mm/s to 199.9 mm/s rms
	Resolution	0.1 mm/s
	Accuracy	± (5% + 2 dgt) (10 Hz to 1 kHz)
Displacement	Range	0.001 mm to 1.999 mm peak-peak (rms 2√2)
	Resolution	0.001 mm
	Accuracy	± (5 % +2 dgt) (10 Hz to 1 kHz)
Piezoelectrical Accelerometer		
Typical frequency response	10 Hz to 15 kHz	
Amplitude Units		
Acceleration	g, m/s <sup>2</sup>	
Velocity	mm/s	
Displacement	mm	
Exteal Piezoelectrical Accelerometer		
DC bias voltage	6 V dc to 7 V dc	
DC bias current	3.5 mA	
Note: Other specifications are the same as those of the inteal piezoelectric accelerometers		

Baery	2 AAA, IEC LR03
Baery life	25 hrs for measurement
Dimensions (L x W x H)	24.1 x 7.1 x 5.8 cm
Weight	220 g
Audio port	Stereo audio output jack (3.5 mm Audio Plug)
Exteal sensor port	M12 plug
<b>Environment</b>	
Operating temperature	-10 °C to 50 °C
Storage temperature	-30 °C to 60 °C
Operating humidity	30% to 90% RH (non-condensing)
Operating altitude	2,000 m
Storage altitude	12,000 m
Ingress Protection	IP 40
Vibration limit	500.0 m/s <sup>2</sup> peak
Drop test	1 m
Electromagnetic Compatibility (EMC)	Inteational IEC 61326-1: Portable, Industrial Electromagnetic Environment; CISPR 11: Group 1, Class A
<p>Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for the inteal function of the equipment itself.</p> <p>Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low-voltage power supply network that supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances. Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.</p>	

## Ordering information



**FLUKE-802EN**

---



**Fluke.** *Keeping your world up and running.®*

**Fluke Corporation**  
PO Box 9090, Everett, WA 98206 U.S.A.

**For more information call:**  
In the U.S.A. (800) 443-5853  
In Canada (800) 36-FLUKE  
From other countries +1 (425) 446-5500  
[www.fluke.com](http://www.fluke.com)

©2025 Fluke Corporation.  
Specifications subject to change without notice.  
03/2025

**Modification of this document is not permitted  
without written permission from Fluke  
Corporation.**