

Data teknis

Fluke TPS VOLTPROBE 400 V



Spesifikasi: Fluke TPS VOLTPROBE 400 V

Specifications	
Range rms	4 V to 680 V
Vnom	10 V
Vmax. contin.	1000 V
Fast Transient Range $V_p < 1 \text{ ms}$	6000
Intrinsic error	0.15%
Operating voltage	600 V CAT IV

Model



Fluke TPS

Fluke. *Memastikan aktivitas Anda terus berjalan dan beroperasi.*

Fluke Corporation

PO Box 9090, Everett, WA 98206 U.S.A.

For more information call:

In the U.S.A. (800) 443-5853

In Europe/M-East/Africa

+31 (0)40 267 5100

In Canada (800)-36-FLUKE

From other countries +1 (425) 446-5500

www.fluke.com

BUT. FLUKE SOUTH EAST ASIA PTE LTD

Menera Satu Sentra Kelapa Gading #06-05

Jl. Bulevar Kelapa Gading Kav. LA# No. 1

Summarecon Kelapa Gading

Jakarta Utara 14240

Indonesia

Tel: +62 21 2938 5922

Email: info.asean@fluke.com

www.fluke.com/id

©2025 Fluke Corporation. Specifications subject to change without notice.

04/2025

Modification of this document is not permitted without written permission from Fluke Corporation.