

**Data teknis** 

# 1586-2586/C | High-Capacity Module without Relay Card, Accredited Calibration



## Ikhtisar Produk: 1586-2586/C | High-Capacity Module without Relay Card, Accredited Calibration

If you want to configure the 1586A yourself to work with the 1586-2586 High-Capacity Module, you will need a relay card. However, if you already have a slot in the back panel of the 1586A configured for use with the 1586-2586, then you don't need another relay card. You can simply slide the 1586-2586 High-Capacity Module into the preconfigured slot and begin taking measurements. The 1586-2586 High-Capacity Module is designed with universal input channels



to measure temperature, resistance, dc voltage, and dc current for data acquisition applications where channel count and scan speed are important.

## Spesifikasi: 1586-2586/C | High-Capacity Module without Relay Card, Accredited Calibration

Specifications	
Maximum Input	50 V
Offset Voltage	<2 μV
3-Wire Inteal Resistance Mismatch	<50 mΩ
Basic CJC Accuracy	0.6 °C



#### Model



1586-2586/C

High-Capacity Module without Relay Card, Accredited Calibration



#### Fluke. Memastikan aktivitas Anda terus berjalan dan beroperasi.

Fluke Corporation

PO Box 9090, Everett, WA 98206 U.S.A.

BUT. FLUKE SOUTH EAST ASIA PTE LTD

Menera Satu Sentra Kelapa Gading #06-05 JI. Bulevar Kelapa Gading Kav. LA# No. 1

Summarecon Kelapa Gading

Jakarta Utara 14240 Indonesia

Tel: +62 21 2938 5922 Email: info.asean@fluke.com

www.fluke.com/id

©2025 Fluke Corporation. Specifications subject to

change without notice.

04/2025

In the U.S.A. (800) 443-5853 In Europe/M-East/Africa +31 (0)40 267 5100 In Canada (800)-36-FLUKE

For more information call:

From other countries +1 (425) 446-5500

www.fluke.com

Modification of this document is not permitted without written permission from Fluke Corporation.

4 Fluke Corporation 1586-2586/C | High-Capacity Module without Relay Card, Accredited Calibration