

テクニカル・データ

Fluke BP291 4800 mAh High Capacity Li-Ion Battery for Fluke 190-Series-II



主な機能

- 4800 mAh rechargeable Li-ion battery
- Compatible with ScopeMeter® 190 Series III, 190 Series II, MDA-500 Series Motor Drive Analyzers and 430-II Series Power Quality Analyzers

製品概要: Fluke BP291 4800 mAh High Capacity Li-Ion Battery for Fluke 190-Series-II

The Fluke BP291 Li-ion Battery is a 4800 mAh high-capacity battery for use in Fluke 190 Series II and Series III ScopeMeter Test Tools, MDA-500 Series III Motor Drive Analyzers and 430-II Series Power Quality Analyzers. BP291 High-Capacity Li-ion batteries include a charge-status indicator so you can quickly and easily verify charge status in the field, and may be charged either in the instrument itself, or using the option [EBC290 External Charger](#).

仕様: Fluke BP291 4800 mAh High Capacity Li-Ion Battery for Fluke

190-Series-II

Specifications*	
Baery	Li-ion, 10.8 V, 5000 mAh, 54 Wh configuration 3ICR19/66-2
Safety	Pollution degree 2
Operating temperature	0 °C to +50 °C (charging may slow or stop at environmental temperatures >40 °C)
Storage temperature	-20 °C to +60 °C
Recommended storage temperature	<35 °C
Humidity	0 °C to 10 °C noncondensing 10 °C to 30 °C @ 95 % (±5 %) rh 30 °C to 40 °C @ 75 % (±5 %) rh
Altitude	3000 m
Weight	330 g
Dimensions	22 mm x 150 mm x 59 mm

モデル



Fluke BP291

4800 mAh High capacity Li-ion battery for Fluke 190 seriesII

Fluke. 動き続ける世界を支える

Fluke Corporation
PO Box 9090, Everett, WA 98206 U.S.A./div>
お問い合わせ先: ©2025 Fluke
フルーク Corporation.
特約店営業部 仕様は、予告なく変更され
TEL : 03-6714-3114 る場合があります。
www.fluke.com/jp/ 04/2025

世界で最も信頼されている
ツール